



CERTIFICATE

< Member No. 4296 >

Company: TÜV Rheinland Vietnam Co., Ltd.

Location of Facility:

**Plot H-1, RF-7E, Thang Long Industrial Park II, My Hao District,
Hung Yen Province, Vietnam**

Registration No.: T-20141

Facility Name: Shielded Room

(Facility for measuring conducted emissions at Telecommunication (Wired) ports)

Registered Instruments:

AAN[1 pair 55 dB, 4 pairs 55 / 65 / 75 dB, Screened type, Coaxial type]

High Impedance Voltage Probe

Impedance converter

***This is to certify that the above measuring facility is
registered in accordance with VCCI Council Rules.***

*** comments on appendix attached**

Date of Registration: April 17, 2023

This Certificate is valid until April 16, 2026

***VCCI Council
Akira Oda***

Akira Oda





TÜV Rheinland Vietnam Co., Ltd. (Member No. 4296)

To: Site Manager

Facility Name: Shielded Room

REGISTRATION No.: T-20141

Comments on Registration

We hereby notify that the Measurement Facility shown in the Certificate is fully examined and approved by the Registration Committee for Measurement Facilities. However, the Committee made the following comment(s) to take into consideration.

We recommend that you calibrate passive (non-electronic) measurement equipment at least once every two years.